

TUV SUD TUV SUD TUV SUD TUV SUD TUV SUD TUV SUD TUV SUD TUV SUD TUV SUD TUV SUD TUV SUD TUV SUD TUV SUD TUV SUD TUV SUD TUV SUD TUV SUD TUV SUD
ZERTIFIKAT ◆ CERTIFICATE ◆ 認證證書 ◆ CERTIFICADO ◆ CERTIFICAT



Product Service

Attestation of Conformity

No. E8A 055975 0081 Rev. 00

Holder of Certificate: **SENTECH KOREA CORP.**
 21-6, Jimokro 75-gil (Sinchon-dong)
 Paju-si, Gyeonggi-do 10880
 REPUBLIC OF KOREA

Name of Object: **Electrical equ. for measurement, control and laboratory use (Breathalyzer)**

Model(s): **ALX3000**

Description of Object: Rated input voltage: 12, 24 V d.c.
 EMC Classification: Class B equipment according to EN 61326-1

Tested according to: EN 55011:2009/A1:2010
 EN 61326-1:2013
 EN 61326-2-3:2013

This Attestation of Conformity is issued on a voluntary basis according to the Directive 2014/30/EU relating to electromagnetic compatibility. It confirms that the listed apparatus complies with all essential requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for testing and certification. See also notes overleaf.

Test report no.: ICRT-TR-E190455-0A

Date, 2019-04-08 
(Byung-Soo Kang)

Page 1 of 1
 After preparation of the necessary technical documentation as well as the EU Declaration of conformity the required CE marking can be affixed on the product. That Declaration of conformity is issued under the sole responsibility of the manufacturer. Other relevant EU-directives have to be observed.
 TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany